



光ディスクからロボット、飛行機まで

# アルス制御

## ALG-500S 検査成績表

S/N : 01AAB28C

検査年月日 : 2020/08/19

検査条件 : 26.8℃ 52%rH

検査入力 (V)	精度 (±mV)	実測値 (V)							
		CH0	CH1	CH2	CH3	CH4	CH5	CH6	CH7
-10.0	20.80	-9.996	-9.996	-9.999	-10.000	-10.000	-10.000	-9.998	-10.000
-9.0	19.82	-8.999	-9.003	-9.000	-8.999	-8.999	-9.001	-9.004	-8.999
-8.0	18.84	-7.998	-7.999	-7.999	-7.998	-8.001	-7.998	-8.002	-8.002
-7.0	17.86	-7.001	-7.005	-7.001	-7.001	-7.002	-7.004	-7.001	-7.001
-6.0	16.88	-6.000	-6.001	-6.000	-6.000	-6.000	-6.000	-6.005	-6.003
-5.0	15.90	-4.997	-4.999	-4.999	-5.003	-5.002	-4.999	-5.002	-4.999
-4.0	14.92	-3.998	-4.002	-3.998	-3.999	-3.998	-3.998	-3.998	-3.998
-3.0	13.94	-3.001	-3.001	-2.997	-2.998	-3.001	-3.001	-3.002	-3.001
-2.0	12.96	-1.995	-2.000	-1.995	-2.000	-1.995	-2.000	-2.000	-1.995
-1.0	11.98	-0.998	-1.003	-0.995	-0.999	-0.999	-0.999	-0.999	-0.999
0.0	11.00	0.002	-0.002	0.002	-0.002	0.007	0.003	-0.002	0.002
1.0	11.98	1.004	1.001	0.999	0.999	1.004	0.999	0.999	1.004
2.0	12.96	2.000	1.999	2.000	2.002	2.000	2.000	2.000	2.000
3.0	13.94	2.998	3.001	2.999	3.001	3.002	3.002	2.996	3.001
4.0	14.92	3.999	3.997	3.998	4.002	3.999	3.998	3.997	3.997
5.0	15.90	4.999	4.999	4.999	5.004	5.004	4.999	4.999	4.999
6.0	16.88	6.000	5.999	5.996	6.000	6.000	6.000	5.995	6.000
7.0	17.86	7.000	6.996	6.997	7.001	6.997	7.001	6.996	6.996
8.0	18.84	7.997	7.997	7.998	7.998	8.001	7.998	7.998	7.998
9.0	19.82	9.001	8.999	8.999	8.999	8.999	8.999	8.994	8.997
10.0	20.80	10.000	9.995	10.000	10.000	10.000	10.000	9.996	9.999
		OK	OK	OK	OK	OK	OK	OK	OK

精度 (±mV) : ALG-500Sの仕様値 (読取值 × (±0.098%) ±11mV) です。